The reliability of supra-patellar transverse sonographic assessment of femoral trochlear cartilage thickness in healthy adults

Roberts, Harry; Moore, Jonathan; Thom, Jeanette

Journal of Ultrasound in Medicine

DOI:

[10.1002/jum.14775](https://doi.org/10.1002/jum.14775)

Published: 01/04/2019

Peer reviewed version

[Cyswllt i'r cyhoeddiad / Link to publication](#)

Dyfyniad o'r fersiwn a gyhoeddwyd / Citation for published version (APA):

Roberts, H., Moore, J., & Thom, J. (2019). The reliability of supra-patellar transverse sonographic assessment of femoral trochlear cartilage thickness in healthy adults. *Journal of Ultrasound in Medicine*, 38(4), 935-946. <https://doi.org/10.1002/jum.14775>

Hawliau Cyffredinol / General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

1 **The reliability of supra-patellar transverse sonographic assessment of femoral trochlear**
2 **cartilage thickness in healthy adults**

3
4 Running title: Femoral cartilage thickness evaluated by sonography

5
6 Harry M. Roberts Ph.D. ^{1,2}, Jonathan P. Moore Ph.D. ¹ & Jeanette M. Thom Ph.D. ^{1,3}

7
8 ¹School of Sport, Health & Exercise Sciences, Bangor University, UK

9 ²School of Biosciences and Medicine, University of Surrey, UK

10 ³School of Medical Sciences, University of New South Wales, Australia

11
12 Corresponding author:

13 Harry Roberts Ph.D.

14 University of Surrey

15 School of Biosciences and Medicine

16 The Leggett Building

17 Daphne Jackson Road

18 Guildford

19 GU2 7WG

20 Phone: +44 (0)1483 688687

21 Email: h.m.roberts@surrey.ac.uk

Abstract

Objectives

To determine the intra-session reliability of femoral cartilage thickness measurements using ultrasonography and extend the pool of normative data for cartilage thickness measurements assessed by ultrasonography.

Methods

77 healthy participants (55 male and 22 female), with an average age of 43 ± 18 (mean $\pm$ SD) years, volunteered. Resting supra-patellar ultrasound was used to image trochlear cartilage thickness on two separate occasions a maximum of 7 days apart. Reliability was evaluated with intraclass correlation coefficients (ICC), Bland & Altman analysis, standard error of measurement (SEM and SEM%), and the smallest real difference (SRD and SRD%). Normative data was assessed using linear, multiple regression models and independent group t-tests.

Results

The test-retest level of agreement at all locations was high (ICC 0.779-0.843), which increased to high-very high in young adults (ICC 0.884-0.920). The SEM% was 8.2-8.3% at all locations and reduced further to 5.4-6.3% in younger adults. The SRD% was between 22.8-23.1% for the full sample and 14.9-17.5% in young adults only. Multiple regression analyses demonstrated that age, weight, female gender and a high physical activity frequency could significantly predict cartilage thickness at all locations ($P < 0.05$); however, female gender was the only significant independent predictor in all models (all $P < 0.01$). Females also had thinner cartilage at all locations ($P < 0.01$).

Conclusion

Supra-patellar ultrasonography demonstrates high intra-tester reliability and measurement precision and is a promising method to assess trochlear cartilage thickness. Being female may impact femoral cartilage thickness more than other potential risk factors for knee osteoarthritis such as age, weight, and high physical activity frequency.

Keywords: Ultrasonography, femoral trochlear cartilage thickness, reproducibility of results, normative data

57 **Introduction**

58 In recent years, ultrasound (US) has been increasingly used to assess cartilage thickness. A commonly adopted
59 technique is axial supra-patellar US imaging (1–6), although longitudinal US scanning of the knee has also been
60 used (7,8). Despite its emergence as a method to assess trochlear cartilage thickness, to date, only a few studies
61 report the validity (through the comparison of US measurements with cadaver specimens or MRI imaging) and
62 reliability of sonographic evaluation of cartilage (7–9). Previous studies using US to measure trochlear cartilage
63 thickness have utilized either a young adult sample (4,5), or have been confined to clinical populations (2,3,10).
64 Therefore, the value of sonographic measurement of trochlear cartilage thickness in a healthy adult sample is
65 restricted by limited normative data. Further examination of the accuracy and repeatability of this technique is
66 required to establish whether US can be used as an effective tool to measure trochlear cartilage thickness.
67 Overall, the ability to reliably measure trochlear cartilage thickness may offer an important tool to for the
68 assessment of patellofemoral disorders, such as patellofemoral pain syndrome, chondromalacia patella and
69 patellofemoral knee arthritis, in both a clinical and research setting.

70

71 The primary aim of this study was to assess the intra-session reliability of cartilage thickness measurements
72 using sonography. Measurement precision was also assessed to identify the smallest change that can be
73 considered actual change and not just a result of a test-re-test error. A secondary aim of this study was to also
74 extend the pool of normative healthy adult data for cartilage thickness measurements assessed by
75 ultrasonography.

76

77 **Materials and methods**

78 Study participants

79 Seventy-seven healthy volunteers (55 male and 22 female), with an average age of 43 ± 18 years, and with an
80 average body mass index (BMI) of 24.9 ± 3.2 , were enrolled. Participants were targeted through word of mouth,
81 poster advertisement, generic emails, and social media from the Bangor University community and the
82 surrounding North Wales area. The inclusion criteria of entry to the study included being: (i) healthy, (ii) male
83 or female, (iii) aged between 18-80 years. Exclusion criteria included: (i) diagnosed osteoarthritis (OA),
84 rheumatoid arthritis, or other inflammatory diseases, (ii) history of knee malalignment (varus / valgus) greater
85 than 15° , (iii) previous knee injury (including meniscus tear or ligament damage or tear), (iv) recent fracture of
86 lower extremity (within last 6 months), (v) current or prior use of lipid-lowering therapy, corticosteroid

injections, or high dose oral steroids (vi) current or past (within last four weeks) glucosamine and/or chondroitin supplementation use, (vii) pregnancy. This study was approved by the local Research Ethics Committee (School of Sport Health and Exercise Sciences (SSHES), Bangor University) and conducted in accordance with the Helsinki Declaration (2013). Written informed consent was obtained from all participants.

Experimental protocol

Participants were required to visit the School of Sport, Health and Exercise Science, Bangor University on two occasions with each session lasting approximately 60 minutes. During the initial visit, participants completed a medical and basic physical activity questionnaire. Anthropometric measurements (body mass and height) were also assessed using a calibrated balance beam scale (SECA, California, USA) and wall-mounted tape measure (SECA, California, USA), respectively. Ultrasonography was subsequently used to obtain images of the femoral articular cartilage as outlined below. All participants completed their first and second visit at the same time of day and within a 7-day period. Participants were also asked to refrain from any strenuous physical activity for 48 hours prior to each visit.

Ultrasonography

The ultrasound (US) assessment was performed using a 12 MHz linear-array probe (Esaote S.P.A. MyLab50 ultrasound, Firenze, Italy) and acoustic coupling gel (Aquasonic 100, Parker Laboratories, Inc, Fairfield, NJ, USA) following a period of between 15-30 minutes of seated rest. With participants lying in a supine position, and with the knee maximally flexed, the superior margin of the patellar was located and a line was marked on the skin using a washable marker at the point immediately above the superior margin of the patellar and at 1 cm intervals in a superior direction. The transducer was placed in a supra-patella transverse position, perpendicular to the bone surface and orientated to optimize the US image (5,9). The location at which the cartilage thickness of the intercondylar notch appeared greatest was marked on the skin and recorded to enable the examiner to return the transducer to the exact location for all subsequent scans. The same researcher performed all ultrasonography scans following training by a consultant rheumatologist with expertise using this technique.

US images were analyzed by 'Image J' software (Image J, National Institute of Health, Bethesda, MD, USA) to determine the minimal cartilage thickness. The distance from the thin hyperechoic line formed at the synovial space-cartilage border to the line formed at the cartilage-bone border was used to measure minimal cartilage

thickness at the lateral facet, medial facet and intercondylar notch. Anatomic reference points used in the present study corresponded to the midpoint of the intercondylar notch and 1 cm apart in the medial and lateral directions were used as an estimate of the cartilage thickness at the medial and lateral facet, respectively (11). Naredo and colleagues previously demonstrated good reproducibility in femoral cartilage thickness measurement when using comparable anatomical reference points (5). Prior to analysis, all images were de-identified by a second researcher for blinded analysis. Based on the pixel resolution (15.8 pixels/mm) of the images captured by ultrasonography, the ImageJ software allowed images to be measured to an accuracy of greater than one-tenth off a mm, or more specifically, one pixel was equal to 0.06 mm. The cartilage thickness of each image was measured in triplicate and an average of the three measurements was used for all data analysis. As required, the image contrast was adjusted to assist in appropriately identifying the hyperechoic line formed at the synovial space-cartilage border to the line formed at the cartilage-bone border.

Statistical analysis

Reliability analysis

Agreement between measurements was evaluated using a one-way mixed, absolute agreement type, intraclass correlation coefficient (ICC) (12). ICC values can be classified as low: 0.20–0.49; moderate: 0.50–0.69; high: 0.70–0.89; or very high: 0.90–1.00 (13). Paired t-tests, together with Bland-Altman plots, were used to provide an indication of the systematic error (14). The standard error of the measurement (SEM) and the SEM% were calculated as previously described (15,16). SEM and the SEM% were used to establish the measurement precision between visit 1 and 2 and to provide a measure of the smallest value that represents a real change in a group of individuals. Furthermore, to calculate the smallest error in a single individual score, the smallest real difference (SRD) and SRD% were also calculated as previously described (17). All analyses were initially completed using the full dataset. Finally, a split-group analysis was performed for each of the following groups: young adults (18-25 years of age), middle-aged adults (26-50 years of age), older adults (≥ 51 years of age), male only, and female only groups. This analysis provided an opportunity to determine whether age or sex of the participant influenced the level of intra-tester reliability and measurement precision.

Analysis of normative cartilage thickness data

Independent t-tests were used to determine differences between measurements made on the right and measurements made on the left knee. If data was not available for either the right or left side (i.e. a measurement

could not be made for one of the locations), both visit 1 and visit 2 data points were removed to ensure an equal sample size. Simple linear regression analyses were performed to determine the relationship between mean cartilage thickness of the right knee (at each location) and participant characteristics (age, body mass, height, and BMI). Multiple linear regression models were subsequently used to explore the relationship between cartilage thickness (at each location) and potential risk factors; including age, BMI, and female gender and high frequency of weekly physical activity. Physical activity was considered 'high' when participants completed structured exercise training on a minimum of 5 days per week. In addition to multiple regression, mean cartilage thickness between sexes was also assessed by creating an equal sized ($n = 17$) sample matched for age and BMI. Independent t-tests were used to determine whether cartilage thickness differences existed between males and females at each measurement location (intercondylar notch, lateral facet, medial facet). For the multiple comparisons of cartilage thickness between the three locations, Bonferroni corrections were used with $P < 0.016$ ($0.05/3$) for statistical significance. Finally, a one-way analysis of variance (ANOVA) was used to assess the differences in cartilage thickness measurements at each measurement location. The left side was not used within the analysis of normative data as the side to side differences were found to be small and within measurement error in the present study. Moreover, others have reported limited side-to-side differences in femoral cartilage thickness measurement (18) and have advocated the use of unilateral OA models in research (19).

Results

The results from the US measurement of cartilage thickness and participant characteristics are displayed in Table 1. A total of 308 knees were scanned (right and left knee of 77 participants on two occasions). This produced a total of 1168 blinded images (77 participants were imaged three to four times per side on two occasions, i.e. visit 1 and visit 2). Some individual images could not be measured as the hyperechoic line formed at the synovial-cartilage border and/or cartilage-bone border could not be clearly delineated. Thus, it was not possible to confidently measure cartilage thickness for 129 (11%), 180 (15%), and 221 (19%) of the available images for the intercondylar notch, medial and lateral facet, respectively. For the cartilage thickness reliability to be assessed a minimum of one image per location was required. Overall, cartilage thickness could be measured in 306 knees (99.4%) at the medial facet, 304 knees (98.7%) at the intercondylar notch, and 296 knees (96.1%) at the lateral facet.

Reliability analysis

The ICC and 95% confidence intervals (95% CI) for the data are shown in Table 2. The ICC's indicate that the level of agreement at all locations was high (ICC between 0.779 – 0.843), with the highest at the intercondylar notch, followed by the medial facet and then the lateral facet. Subsequent analyses revealed that the level of agreement between measurements was considerably improved when considering younger participants (≤ 25 years of age) only (Table 2). In addition, the image quality and clarity were typically better in younger individuals as highlighted in Figure 1. Results also demonstrated that the intra-tester reliability of cartilage thickness measurements was generally similar when male and female groups were analyzed separately (Table 2).

Systemic variation in cartilage thickness measurements at the notch, medial facet, and lateral facet are shown by Bland-Altman plots (Figure 2). The plots suggest that slightly higher variation (i.e. heteroscedasticity) may exist for higher cartilage thickness measurements, particularly at the notch and medial facet. Moreover, results of the paired t-tests showed no significant difference in cartilage thickness between visit 1 and visit 2 for the medial (1.83 vs 1.82 mm; $P = 0.760$) and lateral facet locations (1.81 vs 1.81 mm; $P = 0.860$). However, at the intercondylar notch, a small but significantly greater cartilage thickness measurement was obtained at visit 2, thus indicating that measurements made during the second visit may be systematically higher compared to measurements made at visit 1 (2.03 vs 2.08 mm; $P = 0.016$). When data was split based on the age of the individuals, paired t-tests (visit 1 *versus* visit 2) did not reveal any systematic differences in measurements made in young participants. Furthermore, although mean differences in cartilage thickness measurements tended to be slightly higher in middle-aged and older participants at most measurement locations compared to the young group, a significant difference between visit 1 and visit 2 was only present at the intercondylar notch in the middle-aged participants.

Measurement precision

The SEM is provided for each location in Table 2. The value ranged from 0.15 - 0.17 mm for all locations. In agreement with the intra-class correlation analysis, the SEM was lowest in the split group analysis of young participants only (Table 2). Moreover, the SEM%, which provides a measure independent of units, indicates that differences in groups of individuals above 8.2 - 8.3% can be considered a real change and not the difference associated with measurement error. Overall, the SEM% values for all analyses (Table 2) provide evidence of a relatively low range (5.4 – 9.6%). Moreover, the smallest real change is the measurement error in a single individual cartilage thickness. Table 2 demonstrate that the SRD is between 0.42 and 0.47 mm for all locations.

In relative terms, this equals 22.8 – 23.1%. An improvement in the smallest real change was shown in young participants (0.28 – 0.32 mm) and when analyzing females only (0.27 – 0.34 mm).

Sonographic assessment of cartilage thickness

Femoral cartilage thickness did not differ between the right and left intercondylar notch, or between the right and left medial facet. Although differences were observed at the lateral facet between the left and right knee (1.78 vs 1.88 mm, $P = 0.04$), the difference was small (5.6%) and within the SEM. For this reason (and as previously stated in the methodology), normative data analyses were based on the right side only.

Age was found to have a negative relationship with lateral cartilage thickness in men (Figure 3A). Participant weight was found to have a positive relationship with cartilage thickness at the intercondylar notch (Figure 3C). When females and men were assessed separately, this positive correlation was only found in men (Figure 4C). Participant height was also found to have a positive relationship with cartilage thickness at all locations (Figure 3B); however, when the analysis was separated for males and females, a positive relationship remained between height and lateral facet cartilage thickness in males only. Moreover, a negative relationship between height and intercondylar notch thickness was found in females. In addition, BMI was found to have a negative relationship with lateral and medial facet thickness in men, but not women (Figure 4D). The correlation coefficient, levels of significance and regression equation are presented in Figure 3 for the full dataset, and in Figure 4 for the comparison between males and females.

Age, weight, female gender and high physical activity frequency (> 5 sessions per week) were the independent variables included in the multiple regression model. This four-predictor model was able to account for 28.8% of the variance in femoral cartilage thickness at the intercondylar notch [$F(4, 59) = 5.953$, $P < 0.001$, $R^2 = 0.288$]. However, gender was the only independent variable to significantly contribute to the model. The beta coefficient ($\beta = -0.367$) indicates that in this sample the cartilage thickness in females was 0.38 mm lower than males ($P < 0.01$). At the lateral facet, the regression model could predict 16% of the variance in cartilage thickness [$F(4, 60) = 2.857$, $P < 0.05$, $R^2 = 0.160$]. As per the previous regression model, gender was the only independent variable to significantly contribute to the model ($\beta = -0.253$, $P = 0.008$). Finally, age, weight, female gender and a high physical activity frequency at the medial facet could significantly predict 15.1% of the

variance in cartilage thickness at the medial facet $F(4, 60) = 2.659, P < 0.05, R^2 = 0.151$]. However, again, gender was the only independent variable that contributed to the model ($\beta = -0.355, P = 0.03$).

Analysis of the stable model, i.e. with sex (a dummy variable) as the only predictor variable, the following models were produced: For the femoral intercondylar notch, femoral cartilage thickness in females could be calculated as $2.204 + (-0.450 \times 1) = 1.754$ mm, and in males it could be calculated as $2.204 + (-0.450 \times 0) = 2.204$ mm. The model was significant ($P < 0.001$). For the femoral lateral facet, femoral cartilage thickness in females could be calculated as $1.933 + (-0.219 \times 1) = 1.714$ mm, and in males it could be calculated as $1.933 + (-0.219 \times 0) = 1.933$ mm. This model was also significant ($P < 0.01$). Finally, for the femoral medial facet, femoral cartilage thickness in females could be calculated as $1.946 + (-0.295 \times 1) = 1.651$ mm, and in males it could be calculated as $1.946 + (-0.295 \times 0) = 1.946$ mm. This model was also significant ($P < 0.01$).

Results demonstrated that cartilage thickness was thicker at the intercondylar notch compared to the medial and lateral facet (Figure 5). However, there was no difference in cartilage thickness between the medial and lateral facet ($P > 0.05$). To further assess for differences in mean cartilage thickness between sexes an equal sized ($n = 17$) sample matched for age and BMI was created. Results demonstrated that mean cartilage thickness was lower in females at the intercondylar notch, lateral facet and medial facet than that of the matched male group (Figure 6). The biggest difference in mean cartilage thickness between males and females was at the medial facet (2.00 mm *versus* 1.60 mm, respectively).

Discussion

The use of supra-patellar transverse sonography to assess trochlear cartilage thickness is a novel technique, which required the further study into its reliability and accuracy. The purpose of the present study was to ascertain the intra-tester reliability of supra-patellar transverse US of trochlear cartilage thickness in a group of healthy males and females across a wide range of ages. Notably, the present study demonstrates high intra-tester reliability for trochlear cartilage thickness at the intercondylar notch, medial facet, and lateral facet, as well as a reasonably small measurement error. Additional analysis revealed that both intra-tester reliability and measurement precision reliability was better in young healthy individuals when compared with older counterparts.

In healthy individuals, supra-patellar transverse ultrasonography allowed a quick and straightforward assessment of trochlear cartilage. The high ICCs found in the present study [intercondylar notch 0.843 (0.790 - 0.883), medial facet 0.834 (0.778 - 0.835) and lateral facet 0.779 (0.707 - 0.876)], are comparable to previously reported ICCs using a very similar standardized protocol in a small sample of flexed cadaver knee (age of death was 76-89 years) (9). Interestingly, in both studies, the level of agreement at the lateral facet was lower compared to the intercondylar notch. One possibility is that the lateral and medial facets are prone to an increased level error related to the inclination and positioning of the US transducer (9). This is supported by previous evidence using MRI, which reported that central weight regions often provide greater accuracy than boundary areas (20). Results in the present study also revealed that intra-tester reliability was considerably greater in younger individuals compared to middle-aged and older individuals. Given that a limited degenerative change would be expected in young healthy adults, the increased reliability in young individuals might be due to the trochlear cartilage appearing considerably clearer in young participants. In contrast, image quality in older individuals was often lower, thus reducing the ability of the investigator to delineate images and offer such precise measurements.

Images obtained from this study provided a clear hyperechoic line formed at the synovial-cartilage border and/or cartilage-bone border that allowed femoral cartilage thickness to be assessed in most but not all cases. Compared to the study by Yoon and colleagues, cartilage thickness could be measured in a greater proportion of knees at the medial facet (98.7 vs 70.6%) and lateral facet (96.1 vs 90.1%) in the current study (8). Differences in the ability to measure cartilage thickness between the two studies are likely to relate to the participants (i.e. OA vs healthy individuals in the current study). Several degenerative changes, including, roughened and fibrillated articular cartilage, cartilage loss, asymmetrical narrow, as well as abnormalities at the subchondral bone have previously been associated with poorly defined hyperechoic cartilage borders (5,8). Moreover, despite great care being used to standardize the US assessment of the knee and to replicate the positioning of both the participant and transducer between sessions, other factors such as poor transducer positioning or movement artefact, may also contribute to poor image quality (9). In the current study, of the 7 knees which could not be measured, individuals were all male, mostly older and had a higher BMI. These factors and the relationship with the ability to measure cartilage thickness using US warrant further investigation.

Femoral trochlear cartilage thickness measurements in the present study were comparable to several previous studies using the same US methodology in similarly aged healthy individuals (3,21). In contrast, others have reported slightly greater femoral trochlear cartilage thicknesses in young (25–40 years) healthy individuals compared to the present study (5). The present study also found a significantly thicker cartilage thickness at the intercondylar notch compared to the lateral and medial facets. This difference has not been observed to the same extent in several other studies (3,5,21) and may be related to differences in biomechanical loading. In addition, femoral trochlear cartilage thickness did not differ between the right and left intercondylar notch, or the right and left medial facet. Although differences were observed at the lateral facet between the left and right knee (1.78 vs 1.88 mm, $P = 0.04$), the difference was small (5.6%) and within the SEM. Side to side differences in thickness have previously been reported; however, differences in cartilage thickness tend to be small (total knee joint: $3.8 \pm 3.1\%$) with no significant differences for limb dominance (22). A previous report indicates good correlations between morphological dimensions of the left and right side and advocates the use of unilateral OA models in research (19).

The current study also found females had lower cartilage thickness at all locations compared to males, which is consistent with previous studies using both MRI and US (5,23,24). Furthermore, regression analyses in the present study found that female gender was the only variable that could explain the variation in cartilage thickness. The lower trochlear cartilage thicknesses observed in the present study may relate to differences in body size between men and women. This is supported by the current finding that women have thinner trochlear cartilage thickness compared to men after the adjustment for age and BMI (Figure 6), and previously, after adjustment for body height and weight (24). Differences between males and females may also relate to differences in the sex hormone estrogen (25), which is understood to act upon estrogen receptors found in articular cartilage (26), and/or to differences in the dynamic loading across the knee joint between men and women (27).

Further analyses demonstrated that age was negatively associated with cartilage thickness at the lateral facet, but only in males. Similarly, several studies have previously found ageing to be negatively associated with femoral cartilage thickness assessed by both US (5) and MRI (28). The results of the current study suggest that the lateral femoral facet might be the most prominent site for age-related change. Furthermore, although age has previously been found to be negatively associated with femoral cartilage thickness in both men and women (5),

the present study suggests that men are more at risk of age-related change in cartilage thickness at the lateral facet. This finding is surprising given that older women are at increased risk of OA (29) and may relate to the small sample of females in the present study.

Anthropometric variables such as body height and weight may also influence trochlear cartilage thickness. In the current study, a positive relationship was found between body height and cartilage thickness for all three locations. Several previous studies have also found body height to be positively associated, albeit weakly, with cartilage thickness (24,28). In contrast, a positive relationship between weight and cartilage thickness was observed only at the intercondylar notch. Moreover, when the relationship was explored separately for males and females, weight and body height demonstrated a different relationship with femoral cartilage thickness. Both body height and weight were shown to have a positive relationship with cartilage thickness at various locations in men, while in females, weight was unrelated to femoral cartilage thickness and body height was negatively related at certain locations. Previous research also demonstrated that neither weight nor height was correlated with femoral cartilage thickness in women, and only body height was positively correlated with femoral cartilage thickness in men (30). Reasons for the difference between men and women are unknown. However, it appears that the higher joint loads that are related to body size may have a more favorable impact on cartilage thickness of healthy men compared to women. Whether the relationship extends to a group of men with a greater variation in body size remains unclear. The present study also found that BMI had a negative relationship with both lateral and medial facet thickness in men. This supports previous research indicating having a high BMI may increase the risk of reduced cartilage thickness and knee OA (31). The results of the current study would suggest that while being either heavier or taller may be positive for cartilage thickness in men, an unfavorable body composition may reduce cartilage thickness. This may also suggest that muscle function and physical fitness may have a key role in cartilage thickness morphology. Although exercise frequency as a measure of physical activity level was not associated with cartilage thickness in the present study, future research, together with more refined measures of physical activity is required to explore the potential relationship and determine whether a moderation effect exists.

A primary limitation of this study was the inability to determine the validity of femoral cartilage thickness measurements made using US with a gold standard such as MRI. Nonetheless, US may be regarded as a promising measurement technique that has demonstrated a good agreement in both cartilage thickness

measurements made using US and MRI (7) as well as US and anatomical specimens (9). Nevertheless, this level of agreement is not a universal finding, particularly when using supra-patellar axial US to assess medial facet thickness (8) and when severely damaged knees are included in the analysis (9). Importantly, caution is warranted when considering the validity and reliability of sonographic measures of cartilage thickness when the sample includes older individuals, and individuals with significant knee OA. In addition, unlike the analysis of MRI, US cartilage thickness measurements are largely limited to the femoral plate and do not offer the ability to assess other morphological measurements such as cartilage volume. Unlike MRI, the use of US to detect changes in cartilage thickness following acute loading is unclear. Although acute changes in femoral cartilage thickness following walking and running have recently been reported (32), our recent work demonstrated that such change was not a universal finding (11). A further limitation of the present study relates to the fact that inter-tester reliability was not assessed. This is particularly important given the usefulness of sonographic cartilage thickness measurements as a clinical and research tool relies on the ability to make direct comparisons between studies.

This cross-sectional study of the healthy adults, with a wide age range, demonstrates high intra-tester reliability for all femoral cartilage locations (ICC's between 0.779-0.843) and measurement precision (SEM% between 8.2-8.3%), which is better in younger adults (ICC's between 0.884-0.920 and SEM% 5.4-6.3%). Thus, in younger adults, differences between groups or because of an intervention, that is greater than 6.3% would represent real difference and not just measurement error. Finally, this study also provides normative data for knee cartilage thickness measured by sonography. Considerable variability exists in the femoral cartilage thicknesses of healthy individuals. However, cartilage thickness appears greatest at the intercondylar notch compared to the medial and lateral facets. Furthermore, the data suggest that females have reduced cartilage thickness compared with males and that both ageing and anthropometric measures affect cartilage thickness differently in males and females. This research offers interesting data for the study of the patello-femoral joint.

Acknowledgements

The authors thank all the participants who volunteered for this study.

References

1. Akkaya S, Akkaya N, Ozcakar L, Kilic A, Sahin F, Atalay NS, et al. Ultrasonographic evaluation of the

- 384 femoral cartilage thickness after unilateral arthroscopic partial meniscectomy. *Knee Surg Sports*
 385 *Traumatol Arthrosc.* 2013 May; 21(5):1104–1110.
- 386 2. Kara M, Tiftik TTT, Öken Ö, Akkaya N, Tunc H, Özçakar L, et al. Ultrasonographic measurement of
 387 femoral cartilage thickness in patients with spinal cord injury. *J Rehabil Med.* 2013; 45(2):145–148.
- 388 3. Kaya A, Kara M, Tiftik TTT, Tezcan ME, Ozturk MA, Akinci A, et al. Ultrasonographic evaluation of
 389 the femoral cartilage thickness in patients with systemic lupus erythematosus. *Rheumatol Int.* 2012 Apr;
 390 33(4):899–901.
- 391 4. Malas FÜÜ, Kara M, Aktekin L, Ersoz M, Ozcakar L, Ersöz M, et al. Does vitamin D affect femoral
 392 cartilage thickness? An ultrasonographic study. *Clin Rheumatol.* 2013 Sep; 33(9):1–4.
- 393 5. Özçakar L, Tunç H, Öken Ö, Ünlü Z, Durmu B, Ozcakar L, et al. Femoral cartilage thickness
 394 measurements in healthy individuals: learning, practicing and publishing with TURK-MUSCULUS. *J*
 395 *Back Musculoskelet Rehabil.* 2014;27(2):117–124.
- 396 6. Ozturk GT, Malas FÜÜ, Yildizgoren MT, Baki AE, Inal EE, Batmaz İİİİ, et al. Ultrasonographic
 397 Assessment of the Femoral Cartilage Thickness in Patients with Pes Planus. *Am J Phys Med Rehabil.*
 398 2015 Jul; 94(7):568–572.
- 399 7. Tarhan S, Unlu Z, Goktan C. Magnetic resonance imaging and ultrasonographic evaluation of the
 400 patients with knee osteoarthritis: a comparative study. *Clin Rheumatol.* 2003; 22(3):181–188.
- 401 8. Yoon C-H, Kim H-YH-SH-Y, Ju JH, Jee W-H, Park S-H, Kim H-YH-SH-Y. Validity of the
 402 sonographic longitudinal sagittal image for assessment of the cartilage thickness in the knee
 403 osteoarthritis. *Clin Rheumatol.* 2008;27(12):1507–1516.
- 404 9. Naredo E, Acebes C, Möller I, Canillas F, de Agustín JJ, de Miguel E, et al. Ultrasound validity in the
 405 measurement of knee cartilage thickness. *Ann Rheum Dis.* 2009;68(8):1322–1327.
- 406 10. Akkaya N, Akkaya S, Ozcakar L, Demirkan F, Kiter E, Konukcu S, et al. Ultrasonographic
 407 measurement of the distal femoral cartilage thickness in patients with unilateral transtibial amputation.
 408 *Prosthet Orthot Int.* 2013; 37(4):268–274.
- 409 11. Roberts HM, Moore JP, Griffith-McGeever CL, Fortes MB, Thom JM. The effect of vigorous running
 410 and cycling on serum COMP, lubricin, and femoral cartilage thickness: a pilot study. *Eur J Appl*
 411 *Physiol.* 2016;116(8):1467–1477.
- 412 12. Shrout PE, Fleiss JL. Intraclass correlations: uses in assessing rater reliability. *Psychol Bull.*
 413 1979;86(2):420–428.

- 414 13. Munro B. Correlation. In: Munro B, editor. Statistical methods for health care research. 5th ed.
415 Philadelphia: Lippincott, Williams, and Wilkins; 2005: p. 239–258.
- 416 14. Atkinson G, Nevill A. Statistical methods for assessing measurement error (reliability) in variables
417 relevant to sports medicine. *Sport Med*. 1998;26(4):217–238.
- 418 15. Thomas JR, Nelson KJ. Research methods in physical activity. Champaign (IL): Human Kinetics; 1990.
- 419 16. Beckerman H, Roebroeck ME, Lankhorst GJ, Becher JG, Bezemer PD, Verbeek ALM. Smallest real
420 difference, a link between reproducibility and responsiveness. *Qual Life Res*. 2001;10(7):571–578.
- 421 17. Eliasziw M, Young SL, Woodbury MG, Fryday-Field K. Statistical methodology for the concurrent
422 assessment of interrater and intrarater reliability: using goniometric measurements as an example. *Phys*
423 *Ther*. 1994;74(8):777–788.
- 424 18. Eckstein F, Müller S, Faber SC, Englmeier KH, Reiser M, Putz R. Side differences of knee joint
425 cartilage volume, thickness, and surface area, and correlation with lower limb dominance - An MRI-
426 based study. *Osteoarthr Cartil*. 2002;10(12):914–921.
- 427 19. Dargel J, Feiser J, Gotter M, Pennig D, Koebke J. Side differences in the anatomy of human knee joints.
428 *Knee Surgery, Sport Traumatol Arthrosc*. 2009;17(11):1368–1376.
- 429 20. Koo S, Gold GEE, Andriacchi TPP. Considerations in measuring cartilage thickness using MRI: factors
430 influencing reproducibility and accuracy. *Osteoarthr Cartil*. 2005;13(9):782–789.
- 431 21. Malas FU, Kara M, Kaymak B, Akinci A, Ozcakar L. Ultrasonographic evaluation in symptomatic knee
432 osteoarthritis: clinical and radiological correlation. *Int J Rheum Dis*. 2014;17(5):536–540.
- 433 22. Eckstein F, Faber S, Mühlbauer R, Hohe J, Englmeier K-H, Reiser M, et al. Functional adaptation of
434 human joints to mechanical stimuli. *Osteoarthr Cartil*. 2002;10(1):44–50.
- 435 23. Faber SC, Eckstein F, Lukasz S, Mühlbauer R, Hohe J, Englmeier KH, et al. Gender differences in knee
436 joint cartilage thickness, volume and articular surface areas: assessment with quantitative three-
437 dimensional MR imaging. *Skeletal Radiol*. 2001;30(3):144–150.
- 438 24. Otterness IG, Eckstein F. Women have thinner cartilage and smaller joint surfaces than men after
439 adjustment for body height and weight. *Osteoarthr Cartil*. 2007;15(6):666–672.
- 440 25. Ben-Hur H, Thole HH, Mashiah A, Insler V, Berman V, Shezen E, et al. Estrogen, progesterone and
441 testosterone receptors in human fetal cartilaginous tissue: immunohistochemical studies. *Calcif Tissue*
442 *Int*. 1997;60:520–560.
- 443 26. Ushiyama T, Ueyama H, Inoue K, Ohkubo I, Hukuda S. Expression of genes for estrogen receptors

- alpha and beta in human articular chondrocytes. *Osteoarthr Cartil.* 1999;7:560–566.
27. Cicuttini F, Forbes A, Morris K, Darling S, Bailey M, Stuckey S. Gender differences in knee cartilage volume as measured by magnetic resonance imaging. *Osteoarthr Cartil.* 1999;7(3):265–271.
28. Hudelmaier M, Glaser C, Hohe J, Englmeier KH, Reiser M, Putz R, et al. Age-related changes in the morphology and deformational behavior of knee joint cartilage. *Arthritis Rheum.* 2001;44(11):2556–2561.
29. Felson DT, Naimark a, Anderson J, Kazis L, Castelli W, Meenan RF. The prevalence of knee osteoarthritis in the elderly. The Framingham Osteoarthritis Study. *Arthritis Rheum.* 1987;30(8):914–918.
30. Hudelmaier M, Glaser C, Englmeier K-H, Reiser M, Putz R, Eckstein F. Correlation of knee-joint cartilage morphology with muscle cross-sectional areas vs. anthropometric variables. *Anat Rec A Discov Mol Cell Evol Biol.* 2003;270(2):175–184.
31. Reijman M, Pols HA, Bergink AP, Hazes JM, Belo JN, Lieveense AM, et al. Body mass index associated with onset and progression of osteoarthritis of the knee but not of the hip: the Rotterdam Study. *Ann Rheum Dis.* 2007;66(2):158–162.
32. Harkey MS, Blackburn JT, Davis H, Sierra- Arévalo L, Nissman D, Pietrosimone B. Ultrasonographic Assessment of Medial Femoral Cartilage Deformation Acutely Following Walking and Running. *Osteoarthr Cartil.* 2017; 25(6):907-913.

474
475
476

477 **Table** Error! No text of specified style in document. Physical characteristics of participants and knee cartilage thickness

		Men (n = 55)		Women (n = 22)		Total (n = 77)	
		Mean (SD)	Range	Mean (SD)	Range	Mean (SD)	Range
Age (years)		45 (18)	18 - 70	38 (20)	20 - 79	43 (18)	18- 79
Body height (m)		1.77 (0.06)	1.64 - 1.95	1.67 (0.06)	1.51 - 1.78	1.74 (0.08)	1.5 – 1.95
Body mass (kg)		79.7 (11.1)	63.3 - 120.7	66.3 (11.6)	40.5 – 89.4	75.8 (12.7)	40.5 - 120.7
BMI		25.4 (3.0)	21.0-35.7	23.7 (3.5)	17.7 - 30.2	24.9 (3.2)	17.7 - 35.7
Knee cartilage thickness (mm)							
Right							
	Lateral	1.93 (0.29)	1.43 - 2.73	1.71 (0.29)	1.23 - 2.37	1.87 (0.30)	1.23 - 2.73
	Notch	2.20 (0.40)	1.28 – 3.22	1.75 (0.23)	1.27- 2.36	2.07 (0.43)	1.27 – 3.22
	Medial	1.95 (0.38)	1.15 - 2.97	1.65 (0.28)	1.06 - 2.30	1.86 (0.38)	1.06 - 2.97
Left							
	Lateral	1.79 (0.30)	1.02 - 2.37	1.68 (0.34)	0.93 – 2.44	1.76 (0.32)	0.93 - 2.44
	Notch	2.17 (0.40)	1.44 - 3.12	1.84 (0.32)	1.18 - 2.46	2.08 (0.41)	1.18 - 3.12
	Medial	1.84 (0.34)	1.01 - 2.60	1.63 (0.31)	1.08 – 2.28	1.78 (0.34)	1.01 - 2.60

Table 2 Reliability of cartilage thickness measurements made at visit 1 and visit 2 for all locations with comparisons between age and gender

Location	ICC (95% CI)	SEM	SEM%	SRD	SRD%
Overall (n=77)					
Notch	0.843 (0.790 - 0.883)	0.17	8.2	0.47	22.9
Medial	0.834 (0.778 - 0.876)	0.15	8.2	0.42	22.8
Lateral	0.779 (0.707 - 0.835)	0.15	8.3	0.42	23.1
Young adults (n=20)					
Notch	0.920 (0.854 - 0.957)	0.12	5.7	0.32	15.8
Medial	0.884 (0.792 - 0.937)	0.11	6.3	0.32	17.5
Lateral	0.906 (0.830 - 0.949)	0.10	5.4	0.28	14.9
Middle-aged adults (n=29)					
Notch	0.843 (0.747 - 0.905)	0.18	8.4	0.49	23.3
Medial	0.800 (0.684 - 0.877)	0.13	6.8	0.35	18.9
Lateral	0.639 (0.453 - 0.772)	0.17	9.2	0.47	25.4
Older adults (n=28)					
Notch	0.779 (0.651 - 0.864)	0.19	9.2	0.51	25.4
Medial	0.832 (0.731 - 0.898)	0.17	9.6	0.48	26.5
Lateral	0.788 (0.661 - 0.872)	0.15	8.7	0.42	24.2
Male only (n=55)					
Notch	0.804 (0.725 - 0.862)	0.18	8.3	0.50	23.3
Medial	0.803 (0.725 - 0.861)	0.16	8.7	0.45	24
Lateral	0.744 (0.645 - 0.819)	0.15	8.2	0.42	22.8
Female only (n=22)					
Notch	0.838 (0.723 - 0.908)	0.12	6.9	0.34	19.0
Medial	0.870(0.775 - 0.927)	0.10	5.8	0.27	16.2
Lateral	0.828 (0.708 - 0.902)	0.12	6.8	0.32	19.0

ICC = intra-class correlation; CI = confidence intervals; SEM = standard error of measurement; SRD = smallest real difference); young (≤ 25 years of age), middle-aged (26-50 years of age), and old age groups (≥ 51 years of age)

Figure 1 US transverse image of the femoral articular cartilage demonstrating the difference in image quality and clarity between young, middle-aged, and old groups. Image A) represents the ‘young’ group (23-year-old male), image B) represents the ‘middle-aged’ group (44-year-old male), and image C) represent the ‘old’ group (69-year-old male). M = the location of medial facet; N = the intercondylar notch; L = the lateral facet

Figure 2 The Bland-Altman plots demonstrate the mean difference between the cartilage thickness measurements at visit 2 and visit 1 (i.e. visit 2 minus visit 1) plotted against the mean of the two visits (i.e. visit 1 plus visit 2, divided by 2). Plot A) represents intercondylar notch, B), lateral facet and C) medial facet (solid line represents mean difference and dashed lines represent upper and lower limits of agreement)

Figure 3 Variation of mean femoral cartilage thickness at the intercondylar notch, lateral facet and medial facet with physical characteristics of the participants. A) Age, B) Height, C) Weight, and D) BMI. R-value, significance value, and regression equation are also presented above with significant findings highlighted in bold. Solid, dashed and round dot trendline = intercondylar notch, lateral facet, and medial facet, respectively

Figure 4 Presents variation of mean femoral cartilage thickness at the intercondylar notch, lateral facet and medial facet with physical characteristics of the participants for both males and females A) age, B) height, C) Weight, and D) BMI. R-value, significance value, and regression equation are also presented above with significance highlighted in bold. Black trendline = male; grey trendline = female; Solid line = notch; dashed trendline = lateral; round dot trendline = medial

Figure 5 Mean cartilage thickness measurements at the medial facet, intercondylar notch and the lateral facet: * = significant difference between groups at $P < 0.01$ level. Data are means $\pm$ SD

Figure 6 Mean cartilage thickness measurements at the medial facet, intercondylar notch and the lateral facet for both male ($n = 17$) and female ($n = 17$) participants, matched for age and BMI. * = significant difference between groups at $P < 0.01$ level. Data are means $\pm$ SD

Figure 1. US transverse image of the femoral articular cartilage demonstrating the difference in image quality and clarity between young, middle-aged, and old groups. Image A) represents the 'young' group (23-year-old male), image B) represents the 'middle-aged' group (44-year-old male), and image C) represent the 'old' group (69-year-old male). M = the location of medial facet; N = the intercondylar notch; L = the lateral facet

Figure 2. The Bland-Altman plots demonstrate the mean difference between the cartilage thickness measurements at visit 2 and visit 1 (i.e. visit 2 minus visit 1) plotted against the mean of the two visits (i.e. visit 1 plus visit 2, divided by 2). Plot A) represents intercondylar notch, B), lateral facet and C) medial facet (solid line represents mean difference and dashed lines represent upper and lower limits of agreement)

Figure 3. Variation of mean femoral cartilage thickness at the intercondylar notch, lateral facet and medial facet with physical characteristics of the participants. A) Age, B) Height, C) Weight, and D) BMI. R-value, significance value, and regression equation are also presented above with significant findings highlighted in bold. Solid, dashed and round dot trendline = intercondylar notch, lateral facet, and medial facet, respectively

Figure 4. Presents variation of mean femoral cartilage thickness at the intercondylar notch, lateral facet and medial facet with physical characteristics of the participants for both males and females A) age, B) height, C) Weight, and D) BMI. R-value, significance value, and regression equation are also presented above with significance highlighted in bold. Black trendline = male; grey trendline = female; Solid line = notch; dashed trendline = lateral; round dot trendline = medial

Figure 5. Mean cartilage thickness measurements at the medial facet, intercondylar notch and the lateral facet: * = significant difference between groups at P < 0.01 level. Data are means ± SD

Figure 6. Mean cartilage thickness measurements at the medial facet, intercondylar notch and the lateral facet for both male (n = 17) and female (n = 17) participants, matched for age and BMI. * = significant difference between groups at P < 0.01 level. Data are means $\pm$ SD